



An ISO 9001 : 2008 Port
ISPS CODE Compliant

मुरगांवपत्तनन्यास-चिकित्साविभाग

MORMUGAO PORT TRUST-MEDICAL DEPARTMENT

MEDICAL PROCUREMENT & STORES

HEADLAND, SADA, MORMUGAO, GOA-403 804

Phone: CMO: 0832-2521377, 2594901,

Sr.Dy.MM- 2594918, Stores-2594924/37

E-Mail: sateesh.vernekar@mptgoa.gov.in

balaji.keni@mptgoa.gov.in

Website: www.mptgoa.gov.in

All correspondences should be addressed to the
Chief Medical Officer

NOTICE INVITING LIMITED TENDER

Tender No.	MD/Stores/202	Tender Closing Date & time	23-03-2020 at 16.00 hrs.
-------------------	----------------------	---------------------------------------	---------------------------------

We intend to procure the following materials as detailed below:

P.R.No: Non stock indent		Our Code: N.S.
Sr. No	Item Description	Quantity
1.	Supply of Clinical/Diagnostic Audiometer as per Detailed Specifications at Annexure-I	1 No.
2.	Calibration Charges for on-site (at MPT Hospital) calibration after warranty period of 05years. (Terms & Conditions as per Annexure-II)	
Delivery Period:		4 -6 weeks.


The bidder should submit the Tender in the sealed cover and only in the prescribed format given overleaf, super-scribing the envelope with tender number and opening date and should reach to the office of Chief Medical officer OR Alternately, the offer can also be deposited personally in the Tender Box located in the office of undersigned on or before the closing date/time. The tenders received on or before the closing date & time will be opened on the next working day at 10:00 hrs in the cabin of undersigned.

The terms & conditions applicable are given overleaf. The bidders should accept all the tender terms & conditions and adhere to our delivery period. The Bidders may witness the tender opening if they desire so.

The Administration reserves the right to accept or reject any tender in whole or in part without assigning any reasons thereof.

Thanking You,

भवदीय,


(Nimish Pillai)
Chief Medical Officer

मुरगांवपत्तनन्यास-चिकित्साविभाग

Dispatch No: MD/Stores/ 1879

Date: 5.03.2020

By: Mail/Speed Post/ Ordinary Post.

To,
M/s. Elkon Pvt Ltd; Mumbai

M/s. Global Systems Pvt Ltd; Marcel, Ponda,Goa

M/s. Audest Pvt Ltd

M/s. Ansan Pvt. Ltd; Margao, Goa

M/s. Arphi Electronics Pvt. Ltd: Mumbai

Website

Sl. No	Item Description	Quantity
1.	Supply of Clinical Diagnostic Audiometer as per Detailed Specifications of Annexure-I	1 No.
2.	Calibration Charges for on-site (at MPT Hospital) calibration after warranty period of 05 years.	
	(Terms & Conditions as per Annexure-II)	
Delivery Period: 4-6 weeks.		

The bidder should submit the Tender in the sealed cover and only in the prescribed format given overest, super-scribing the envelope with tender number and opening date and should reach to the office of Chief Medical Officer OR Alternatively, the offer can also be deposited personally in the Tender Box located in the office of undersigned on or before the closing date. The tenders received on or before the closing date & time will be opened on the next working day at 10:00 hrs in the cabin of undersigned.

The terms & conditions applicable are given overest. The bidders should accept all the tender terms & conditions and adhere to our delivery period. The Bidders may witness the tender opening if they desire so.

The Administration reserves the right to accept or reject any tender in whole or in part without assigning any reasons therefor.

Thanking You,

भवदीय

(Ministry Officer)
Chief Medical Officer

Annexure-I

SPECIFICATIONS FOR CLINICAL /DIAGNOSTIC AUDIOMETER

1. CHANNELS: Two separate channels one for Stimulus and one for masking.
2. FREQUENCIES: Air: 250-500-1000-1500-2000-3000-4000-6000-8000 Hz.
Bone: 250-500-1000-1500-2000-3000-4000 Hz.
3. ATTENUATOR RANGE: -10 to 120 dB steps with range extender function.
4. HEARING LEVEL RANGE: Values for Air: Max 120 dB HL at 500-6000 Hz.
Max 110dB at 250Hz
Max 100 dB at 8000Hz.
Values for Bone: 70 dB HL at 500-4000 Hz.
60dB at 250Hz.
5. OUTPUTS:
 - Phones: Air Conduction
 - Bone: Bone Conduction
 - F.F: Free field loudspeaker via external amplifier.
 - Output voltage: max. 1.5 Vrms
6. MASKING: Wide Band
7. PULSING RATE: 1.5 Hz
8. TONE: Continuous, Pulsed
9. DISPLAY: Alphanumeric transreflective LCD. LED backlighting.
10. POWER SUPPLY: 230V AC 50 Hz.
11. Data Storage: Audiometric test results to be saved in the PC or printed using printer.
12. STANDARD ACCESSORIES:
 - Head set
 - Bone Conductor with headband
 - Patient response hand switch.
 - Dust Cover
 - Operation Manual


Chief Medical Officer

TENDER TERMS AND CONDITIONS

1. **PRICES:** The bidders should quote FIRM prices and should be on F.O.R. destination basis to be supplied to our stores at MPT Hospital on door delivery basis.
2. **DELIVERY:** The instrument shall be supplied within 4-6 weeks from the date of receipt of purchase order.
3. The instrument should cover the Guarantee/Warranty period as per the Manufacturer's policy.
4. **Performance Guarantee/Retension Money:** 10% of the invoice value (before GST) will be retained from the bill & the retained amount will be released on completion of a defect liability period.
5. **PAYMENT:** Our payment terms will be 100% within 30 days on receipt of materials/bills whichever is later only through EFT.
6. **LIQUIDATED DAMAGES:** In the event of delay in supply, liquidated damages at 1% of the total basic value of the undelivered quantity will be levied per week or part thereof of delay. The total liquidated damages shall however not exceed 10% of the total basic value of purchase Order. The total basic value here means the total value of the Order exclusive of GST, etc. The date of receipt of the item in our Stores will be considered as date of delivery for the purpose of liquidated damages.
7. **FORCE MAJEURE & EXTENSION IN DELIVERY PERIOD:** In case of failure to deliver the material/medicine in time which shall have arisen from war, insurrection, restraint imposed by the Govt., Act or legislation or other authority, accident, strike, riot, lockout or unforeseen events beyond human control directly or indirectly interfering with the supplies of stores or from any cause which the purchaser may admit as reasonable ground for an extension of time, the purchaser will allow such additional time he considers reasonable provided that the contractors/suppliers shall report immediately to the purchaser on the occurrence of any such cause as aforesaid which will or might affect the performance of the contract. Request for extension received after expiry of the delivery schedule will not be entertained.
8. Kindly submit catalogue/technical leaflet along with the offer.
9. **Calibration:** The instrument supplied is required to be calibrated at on-site (MPT Hospital, Sada) once the warranty period is over. You are therefore requested to quote for the calibration charges to be carried on-site at MPT hospital, Headland Sada for a period of 05 years after the Warranty /Guarantee Period is over.


Chief Medical Officer